

**URGENT RECALL**

Friday, August 12, 2016

Dear Valued Life-Assist Customer,

According to our records you may have purchased an item that has been recalled by the manufacturer. Please examine your stock to determine if you have the following product, with an affected lot number, in your possession.

**Toomey Syringe, 70cc**

Life-Assist, Inc. Product Code	NDC/UPC	Lot #(s)
OK38	0038460	NGZC3768, NGZC5029, NGZD0924, NGZD1664, NGZD1983, NGZD3103, NGZD3241, NGZD3989, NGZD4053, NGZD4923, NGZE1190, NGZE2134, NGZE3406, NGZF0305, NGZF1546, NGZF2589, NGZF3934, NGZF4488, NGZF4885

Reason for Recall: The manufacturer, C.R. Bard, Inc., is recalling specific product codes and lot numbers that may be at risk for having a slit defect on the packaging which could potentially compromise the product's sterility.

Contact Life-Assist Customer Service at 800-824-6016 or [saleservice@life-assist.com](mailto:saleservice@life-assist.com) for replacement or credit and to obtain a Return Authorization.

**Life-Assist, Inc. cannot replace or provide credit for any product returned that is not impacted by this voluntary action. Please check the product's lot number to verify that it is affected.**

We apologize for any inconvenience.

